



MICRON SCALE RULER WITH LOUPE USER MANUAL



Calibration ISO 17025:2017

ISO 9001
BUREAU VERITAS
Certification



ISO 9001:2015




Item No	Range		Reso- lution	Accu- racy	Vision	Protection	Preset	Display
	mm	inch	mm	µm				
142505151	0-500	0-20"	0,001	±20	10x magnifier	IP67	•	Swiss electronics
142510151	0-1000	0-40"		±30				
142515151	0-1500	0-60"		±40				
142520151	0-2000	0-80"		±50				
142530151	0-3000	0-120"		±70				

BUTTON FUNCTIONS



SET DATA 0



OPERATION INSTRUCTIONS



MICROTECH

Wipe with a clean cloth, soaked in gasoline, bases and measuring surface of the frame and gauge calipers to remove anti-corrosion oil. Then wipe them with a clean dry cloth.

If necessary, open the battery cover; insert the battery (type CR2032) according to the polarity of the electrodes.

This caliper has Autoswitch on/ off function:

- move electronic module for switch on caliper
- after 10 minutes without any moving caliper will switch off

During the measurement, measuring jaws should to sum to the measured object without knocking.

During the measurement avoid warps of measuring surfaces of the instrument. Measuring surface must be fully in contact with the measurement object

WARNING!

IN THE PROCESS OF WORKING WITH SCALES SHOULD BE AVOIDED:

- Scratches on the measuring surfaces and bases;**
- Measuring the size of object in the process of machining;**
- Shocks or dropping, avoid bending of rod or other surfaces.**



LOUPE 10X WITH LED

Micron scale equipped with 10x loupe with LED light. Switch on LED light by screw rotating



ROTATE



MICROTECH



MICROTECH

innovative measuring instruments

61001, Kharkiv, Ukraine, str. Rustaveli, 39

www.microtech.ua
tool@microtech.ua

Change without prior notice

Edition: M1030_0520